# Sucrose Synthase (SS-I, Breakdown Direction) Activity Assay Kit

**Note:** It is necessary to predict 2-3 large difference samples before the formal determination.

**Operation Equipment:** Spectrophotometer

**Cat No:** NA0318 **Size:** 50T/24S

## **Components:**

Extract solution: Liquid 30 mL×1, store at 4°C;

Reagent I: Liquid 8 mL×1, store at 4°C;

Reagent II: Powder×1, store at -20°C and protect from light. Add 2.5 mL Reagent I when the solution will be used. Mix thoroughly. The rest of reagent store at -20°C; Do not freeze and thaw repeatedly;

Reagent III: Liquid 8 mL×1, store at 4°C and protect from light;

Standard: Powder×1, 20 mg fructose. Store at 4°C. Add 1 mL distilled water when the solution will be used and make the final concentration is 20 mg/mL fructose solution for standby. Mix thoroughly. The rest of reagent store at 4°C for one week;

# **Product Description:**

Sucrose synthetase (SS) is the key enzyme in the sugar metabolism of plants. It is responsible for the reversible reaction of sucrose decomposition and synthesis. Its decomposition activity can catalyze the hydrolysis of sucrose to UDPG and fructose. It is also involved in the synthesis of starch, cellulose and hemicellulose.

SS-I can catalyze the production of fructose and UDPG from sucrose and UDP. Fructose reacts with 3,5-Dinitrosalicylic acid to form brownish red substance with characteristic absorption peak at 540 nm. The activity of SS-I can be calculated by measuring the change of absorption value at 540 nm.

#### **Required but Not Provided:**

Spectrophotometer, desk centrifuge, water-bath, transferpettor, 1 mL glass cuvette, mortar/homogenizer, EP tube, ice and distilled water.

## **Protocol**

#### I. Preparation:

Tissue: according to the ratio of mass (g): extraction volume (mL): 1:5-10 to add the extract solution. It is suggested that add 1 mL of extract solution to 0.1 g of tissue. Homogenize on ice. Centrifuge at 8000 g 4°C for 10 min. Take the supernatant on ice for test.

#### **II. Determination procedure:**

- 1. Preheat spectrophotometer/microplate reader for 30 min, adjust wavelength to 540 nm, set the counter to zero with distilled water.
- 2. Dilute 20 mg/mL standard solution with distilled water to 5, 4, 3, 2, 1 mg/mL standard solution for

# standby.

## 3. Operation table:

Control tube (A <sub>C</sub> )	Test tube (A <sub>T</sub> )	Standard tube (A <sub>S</sub> )	Blank tube (A <sub>B</sub> )
20	20	-	-
-	-	20	-
1	-	-	20
80		80	80
-	80	-	-
Mix thoroughly. Put it in 30°C-water bath for 30 min, and in			
95°C water bath for 10 min (cover tightly to prevent water			
100	100	100	100
Mix thoroughly. Put it in boiling water for 5 min (cover tightly to prevent water loss). Cool to room			
temperature.			
800	800	800	800
	20  - 80 - it in 30°C-water bath 100 it in boiling water	20 20	20 20 -  - 20  - 80 80  - 80 -  it in 30°C-water bath for 30 min, and in 10 min (cover tightly to prevent water  100 100 100  it in boiling water for 5 min (cover tightly to prevent water

Mix thoroughly. Add 1 mL reaction solution to the 1 mL glass cuvette. Determine the absorption value A at 540 nm. Record as,  $A_{C_1}A_{T_2}A_{S_3}$ ,  $A_{B_1}$ ,  $\Delta A = A_{T_1}A_{C_2}$ .  $\Delta A_{S_2} = A_{S_3}A_{S_4}$ . Each test tube should be provided with a contrast tube. The standard curve only needs to test once or twice.

## **III. SS-I Calculation:**

#### 1. Standard curve

Take the concentration of each standard solution as x-axis, and the corresponding  $\Delta A$  standard is y-axis. Then the linear regression equation y=kx+b is obtained. Bring  $\Delta A$  into the equation to get x (mg/mL).

#### 2. Calculation

# 1) Protein concentration:

Unit definition: One unit of enzyme is defined as the amount of enzyme that catalyzes the decomposition the consumes of 1 µg fructose per minute every mg tissue protein in the reaction system.

SS-I (U/mg prot) = 
$$x \times V_{SA} \div (Cpr \times V_{SA}) \div T \times 10^3 = 33.33x \div Cpr$$

# 2) Sample weight:

Unit definition: One unit of enzyme is defined as the amount of enzyme that catalyzes the decomposition the consumes of 1  $\mu$ g fructose per minute every gram tissue weight in the reaction system.

SS-I (U/g weight) =
$$x \times V_{SA} \div W \div T \times 10^3 = 33.33x \div W$$

V<sub>SA</sub>: Extract solution volume of cells, 1 mL;

10<sup>3</sup>: Unit conversion factor, 1 mg =  $10^3$  µg;

T: Reaction time, 30 min;

Cpr: Protein concentration, mg/mL;

W: Sample weight, g.

#### Note:

- 1. When the A or  $\Delta A$  is greater than 1, it is recommended to measure after dilution. The dilution ratio needs to be multiplied in the calculation formula.
- 2. Cover the EP tube tightly at 95°C water bath to prevent water loss. It needs to be cooled to room temperature before the next operation. It is to avoid splashing and scalding of liquid and to avoid affecting the test data.

## **Experimental examples:**

1. Take 0.1 g of ryegrass and add 1 mL of Extract solution for sample processing. Follow the measurement procedure after taking the supernatant. Calculate  $\Delta A = A_T - A_C = 0.667 - 0.128 = 0.539$ . Bring the result into the standard curve y=0.1981x-0.0491, and calculate x=2.9687. The enzyme activity is calculated according to the sample mass.

SS-I (U/g weight) = $33.33x \div W = 989.4677 U/g weight$ .

# **Related products:**

NA0823/NA0581 Sucrose Synthetase(SS) Activity Assay Kit
NA0821/NA0579 Sucrose Phosphoric Acid Synthetase(SPS) Activity Assay
NA0582/NA0824 Neutral Invertase(NI) Activity Assay Kit
NA0694/NA0453 Plant Sucrose Content Assay Kit