

Glutaminase (GLS) Assay Kit

Note: Take two or three different samples for prediction before test.

Operation Equipment: Spectrophotometer/ microplate reader

Catalog Number: NA0512

Size:100T/48S

Components:

Extract solution: 70 mL ×1, storage at 4°C. Preheat at 37°C before use.

Reagent 1: powder×1, storage at 4°C. Add 5 mL of Extract solution to dissolve the reagent before use.

Reagent 2A: 0.4 mL×1, storage at 4°C.

Reagent 2B: 1.6mL×1, storage at 4°C. Before use, pour Reagent 2A into Reagent 2B to mix (A:B=1:4 ratio), or prepared when the solution will be used according to the volume ratio Reagent 2A : Reagent 2 B = 1:4 before use.

Reagent 3: 2 mL×1, storage at room temperature.

Standard: 1mL ×1, storage at 4°C. 10 μmol/mL nitrogen standard solution. Preheat at 37°C before use.

Product Description:

GLS (EC3.5.1.2) is mainly found in higher animals, some bacteria and plant roots, catalyze the hydrolysis of glutamine into glutamic acid and ammonia, which plays an important role in the regulation of nitrogen metabolism, especially the regulation of free ammonia and urea metabolism.

The kit uses the indophenol blue colorimetric method to determine ammonia produced by glutamine of GLS-catalyzed to indicate activity.

Reagents and Equipment Required but Not Provided:

Spectrophotometer/microplate reader, adjustable pipette, balance, mortar/homogenizer, centrifuge, micro glass cuvette/96 well flat-bottom plate, ice and distilled water.

Sample preparation:

1. Tissue: The mass (g): Extract solution volume (mL)= 1:5-10, suggested 0.1g of tissue, add 1 mL of Extract solution. After homogenizing on ice, centrifuge 12000g at 4 ° C for 15 min, then take supernatant to be tested.
2. Bacteria or cells
Accordance ratio bacteria or cell amount (10^4): volume of Extract solution(mL)=500~1000:1. Suggested 5 million with 1 mL of Extract solution. Use ultrasonic to splitting bacteria or cell (placed on ice, powder: 300W, work time 3s, interval 7s, total time 3 min). Centrifuge at 12000g at 4 ° C for 15 min. then take supernatant on ice to be tested.

Procedure:

1. Preheat Spectrophotometer/microplate reader for 30 min, adjust the wavelength to 630 nm and set the

counter to zero with distilled water.

- Dilute the standard solution 8 times with the Extract solution to obtain the standard solution of 1.25 $\mu\text{mol/mL}$.
- Sample determination:

Reagent name (mL)	Test tube (At)	Control tube (Ac)	Standard tube (As)	Blank tube (Ab)
Sample	16	16	-	-
Extract	-	64	-	80
Reagent 1	64	-	-	-
Mix and react for 60 min at 37°C			-	-
Standard	-	-	80	-
Reagent 2	16	16	16	16
Reagent 3	12	12	12	12
Distilled water	92	92	92	92

Mix well, react for 30min at room temperature. Measure the absorbance at 630 nm. Recorded as At, Ac, As, Ab. Calculate $\Delta A_s = A_s - A_b$, $\Delta A_t = A_t - A_c$.

Calculation:

- Protein concentration

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the production of 1 μmol of $\text{NH}_3\text{-N}$ per hour every milligram of tissue protein.

$$\text{GLS (U/mg prot)} = \Delta A_t \div (\Delta A_s \div C_{st}) \times V_{er} \div (V_{sa} \times C_{pr}) \div T = 6.25 \times \Delta A_t \div \Delta A_s \div C_{pr}$$

- Fresh weight

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the production of 1 μmol of $\text{NH}_3\text{-N}$ per hour every gram of sample.

$$\text{GLS (U/g)} = \Delta A_t \div (\Delta A_s \div C_{st}) \times V_{er} \div (W \div V_e \times V_{sa}) \div T = 6.25 \times \Delta A_t \div \Delta A_s \div W$$

- Number of cells

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the production of 1 μmol of $\text{NH}_3\text{-N}$ per hour every 5×10^4 cells.

$$\text{GLS (U/mL)} = \Delta A_t \div (\Delta A_s \div C_{st}) \times V_{er} \div (V_{sa} \div V_e) \div T = 6.25 \times \Delta A_t \div \Delta A_s$$

Cst: Standard solution concentration, 1.25 $\mu\text{mol/mL}$;

Vsa: sample volume added, 0.016 mL;

Ver: Volume of enzymatic reaction, 0.08 mL;

Ve: volume used in the extraction process, 1 mL;

Cpr: sample protein concentration, mg/mL;

W: Fresh weight of sample, g;

T: React time, 1 hour.

Note:

- If $\text{OD} > 0.8$, It is recommended to further dilute the supernatant and then measure it.

2. Reagent 2 should be used as soon as possible. If discoloration is found, it can no longer be used.

Recent Product citations:

[1] Fu Y, Lei F, Wang J, et al. Maternal Cigarette Smoke Exposure Disturbs Glutamate/GABA Balance in pFRG of Neonatal Rats[J]. Respiratory Physiology & Neurobiology, 2020: 103383.

[2] Liu S, Li N, Lin Q, et al. Glutaminase 1 in mandarin fish *Siniperca chuatsi*: Molecular characterization, expression pattern and function involving in virus replication[J]. Aquaculture, 2020: 734924.

References:

[1] Mahajan R V, Saran S, Kameswaran K, et al. Efficient production of L-asparaginase from *Bacillus licheniformis* with low-glutaminase activity: optimization, scale up and acrylamide degradation studies[J]. Bioresource technology, 2012, 125: 11-16.

Related Products:

NA0865/NA0622 Nitrate reductase (NR) Activity Assay kit

NA0754/NA0512 Glutaminase (GLS) Assay Kit

NA0751/NA0509 Nitrite Assay Kit (Water And Soil)