

Oxidized Glutathione (GSSG) Content Assay Kit

Note: Take two or three different samples for prediction before test.

Operation Equipment: Spectrophotometer/microplate reader

Catalog Number: NA0539

Size: 100T/96S

Components:

Reagent I: 100 mL×1. Storage at 4°C.

Reagent II: 130 μL×1. Storage at 4°C.

Reagent III: 20 mL×1. Storage at 4°C.

Reagent IV: 2.5 mL×1. Storage at 4°C.

Reagent V: Powder ×1. Storage at 4°C. Dissolve with 2.5 mL of distilled water when the solution will be used, then split into smaller packages, store at -20°C.

Reagent VI: 12.5 μL×1. Storage at 4°C. Prepare Reagent VI and distilled water according to the sample size at the ratio of 1:20 (V: V) before use.

Standard: Powder 10 mg×1. Storage at 4°C.

Product Description

Oxidized Glutathione (GSSG) is an oxidized form of glutathione (GSH), also known as dithioneglutathione, formed by the oxidation of two molecules of glutathione. GSSG is reduced to GSH by glutathione reductase, so most of the body is in the reduced form. The determination of GSH and GSSG content and ratio of GSH/GSSG in cells can reflect the redox status of cells. This kit utilizes reaction of glutathione and 5, 5'-dithiobis-2-nitrobenzoic acid (DTNB) to produce 5-thio-2-nitrobenzoic acid. 5-thio-2-nitrobenzoic acid has the largest absorption at wavelength of 412nm, and 2-Vinylpyridine inhibits reduced glutathione in the original of samples, and then using glutathione reductase to reduce GSSG to GSH, determining the content of Oxidized Glutathione.

Technical Specifications

Minimum Detection Limit: 3.211 μg/mL

Linear Range: 3.9-125 μg/mL

Reagents and Equipment Required but Not Provided.

Analytical balance, mortar/homogenizer, refrigerated centrifuge, water-bath, adjustable pipette, spectrophotometer/ microplate reader, micro glass cuvette/96 well flat-bottom plate.

Procedure

I. Sample preparation

1. Tissue sample

Wash fresh tissues with PBS for twice, then add 0.1 g of sample into homogenizer (the homogenizer has

been rinsed with Reagent I and placed on ice before use). Add 1 mL of Reagent I (the proportion of tissue and reagents can be kept constant), fully grinding on ice (using liquid nitrogen will have a better grinding effect). Centrifuge at $8000 \times g$ and $4^{\circ}C$ for 10 minutes, take the supernatant and place it at $4^{\circ}C$ for test. (The supernatant can be stored at $-80^{\circ}C$ for 10 days.)

2. Blood sample

Plasma: Sample is centrifuged at $600 \times g$ and $4^{\circ}C$ for 10 minutes. Absorbing the upper plasma into another tube add with same volume Reagent I. Centrifuge at $8000 \times g$ and $4^{\circ}C$ for 10 minutes, take the supernatant and place it at $4^{\circ}C$ for test. (The Supernatant can be stored at $-80^{\circ}C$ for 10 days.)

Blood cell: Sample is centrifuged at $600 \times g$ and $4^{\circ}C$ for 10 minutes. Discarding the upper plasma, wash with treble volume of PBS for 3 times (mix blood cell with PBS centrifuge at $600 \times g$ for 10 minutes), add equal volume of Reagent I. After mixing, it is placed at $4^{\circ}C$ for 10 minutes. Centrifuge at $8000 \times g$ for 10 minutes, take the supernatant and place it at $4^{\circ}C$ for test. (The supernatant can be stored at $-80^{\circ}C$ for 10 days.)

3. Cell sample

Harvesting cell should not less than 10^6 , then wash with PBS for twice (mix cell with PBS centrifuge at $600 \times g$ for 10 minutes), mix precipitated cell with the volume of PBS for 3 times. Repeated freezing and thawing 2-3 times (suggest frozen in liquid nitrogen, dissolved in $37^{\circ}C$ water bath). Centrifuge at $8000 \times g$ for 10 minutes, take the supernatant and place it at $4^{\circ}C$ for test. (The supernatant can be stored at $-80^{\circ}C$ for 10 days.)

II. Procedure

1. Preheat spectrophotometer/microplate reader for 30 minutes, adjust the wavelength to 412 nm, set the counter to zero with distilled water.
2. Preheat Reagent II in water bath for 30 minutes: $37^{\circ}C$ (mammal cell) or $25^{\circ}C$ (other species).
3. The standard dilution: dissolve standard with 1 mL of distilled water ($4^{\circ}C$) to concentration of 10 mg/mL. Take suitable solution to prepare the standard of concentration of $125\mu g/mL$ 、 $62.5\mu g/mL$ 、 $31.25\mu g/mL$ 、 $15.625\mu g/mL$ 、 $7.8125\mu g/mL$ 、 $3.90625\mu g/mL$ and $0\mu g/mL$ (The diluent is a ten-fold diluted Reagent I).
4. Add $20 \mu L$ of diluted standard or sample to $0.5 mL$ centrifuge tube, add $1 \mu L$ of Reagent II, incubate at $37^{\circ}C$ for 30 minutes after mixing .
5. Make standard curves

After the incubation, add $140 \mu L$ of Reagent III, $20 \mu L$ of Reagent IV, $20 \mu L$ of Reagent V, and $2 \mu L$ of Reagent VI to the standard tube in sequence. After rapid mixing, the light absorption A_1 and A_2 of 30 s and 150 s respectively were measured at 412 nm. Absorbance ($A_2 - A_1$) is the abscissa (x) and concentration is the ordinate (y), making the standard curve.

Add $140 \mu L$ of Reagent III, $20 \mu L$ of Reagent IV, $20 \mu L$ of Reagent V, and $2 \mu L$ of Reagent VI to the sample tubes in sequence. After rapid mixing, the light absorption A_1 and A_2 of 30 s and 150 s respectively were measured at 412 nm, $\Delta A = A_2 - A_1$.

III. Calculations

According to the standard curve, sample ΔA into the formula (x), calculate the sample concentration of y ($\mu\text{g}/\text{ml}$).

1) Protein concentration

$$\text{GSSH } (\mu\text{g} / \text{mg prot}) = y \times V_{rv} \div V_{rv} \div C_{pr} = y \div C_{pr}$$

2) Sample weight

$$\text{GSSH } (\mu\text{g} / \text{g}) = y \times V_{rv} \div (V_{rv} \div V_{sv} \times W) = y \div W$$

3) Cell amount

$$\text{GSSH } (\mu\text{g} / 10^6 \text{ cell}) = y \times V_{rv} \div (V_{rv} \div V_{sv} \times N) = y \div N$$

4) Solution volume

$$\text{GSSH } (\mu\text{g} / \text{mL}) = y \times V_{rv} / V_s = 2y$$

N: Cell amount, 10^6 ;

V_{rv} : Total reaction volume, 0.203 mL;

V_{sv} : The volume of supernatant was added into the reaction system, $20 \mu\text{L} = 0.02 \text{ mL}$;

W: Sample weight, g;

C_{pr} : Supernatant protein concentration, mg/mL .

Notes:

1. The sample needs to be homogenized completely. If the test cannot be completed temporarily, it can be stored at -80°C .
2. If the content of GSSG content in the sample is uncertain, several gradients can be diluted before measurement.
3. This method uses the enzymatic reaction rate to calculate the substrate concentration and complete readings as accurately as possible at 30 and 150 seconds.
4. Reagent I contained protein precipitant, the supernatant could not be used for protein concentration determination. If the protein content needs to be determined, take another tissue.

Recent Product citations

[1] Ming Song, Fangfang Chen, Yihui Li, et al. rimetazidine restores the positive adaptation to exercise training by mitigating statin-induced skeletal muscle injury. *Journal of Cachexia, Sarcopenia and Muscle*. November 2017; (IF10.754)

[2] Hua Li, Lanying Wang, Yanping Luo. Composition Analysis by UPLC-PDA-ESI (-)-HRMS and Antioxidant Activity Using *Saccharomyces cerevisiae* Model of Herbal Teas and Green Teas from Hainan. *Molecules*. October 2018; (IF3.06)

Reference:

[1] Alpert A J, Gilbert H F. Detection of oxidized and reduced glutathione with a recycling postcolumn reaction[J]. *Analytical biochemistry*, 1985, 144(2): 553-562.

[2] Owens C W I, Belcher R V. A colorimetric micro-method for the determination of glutathione[J].

Biochemical Journal, 1965, 94(3): 705.

Related products:

- NA0781/ NA0540 Reduced Glutathione (GSH) Assay Kit
- NA0779/ NA0538 Glutathione Peroxidase Assay Kit
- NA0839/ NA0597 Glutathione S-transferase(GST) Activity Assay Kit
- NA0782/ NA0541 Glutathione Reductase (GR) Assay Kit