

Hydroxyproline (HYP) Activity Assay Kit

Note: Take two or three different samples for prediction before test.

Detection equipment: Spectrophotometer

Cat No: NA0849

Size: 50T/48S

Components:

Extract solution: 6 mol/L hydrochloric acid (HCl), self-provided reagent. Concentrated HCl (37%): H₂O (V/V) =1:1, stored at room temperature.

Reagent I: 15 mL×1, store at 4°C and protect from light.

Reagent II: 15 mL×1, store at 4°C and protect from light.

Standard: 1 mL×1, 1 mg/mL hydroxyproline, store at 4°C.

Description:

HYP is one of the main components of collagen in the body. Most of the collagen is distributed in the skin, tendon, cartilage and blood vessels et al. Therefore, the content of HYP is an important index reflecting the metabolism and fibrosis degree of collagen tissue.

The sample is hydrolyzed to produce free HYP, which is further oxidized by chloramine T. The oxidized product reacted with p-Dimethylaminobenzaldehyde to produce red compound with characteristic absorption peak at 560 nm. The content of HYP can be calculated by measuring the absorption value of sample hydrolysate at 560 nm.

Required but not provided:

Scales, oven, glass tube, centrifuge, water bath, spectrophotometer, 1 mL glass cuvette, 6 mol/L HCl and distilled water.

Procedure:

I. Sample preparation

1. Tissue sample:

Weigh about 0.2 g of the sample into the glass tube, cut the tissue into pieces as much as possible for digestion. Add 2 mL of Extract solution and the cover is slightly loose and not airtight, boil it or bake it in 110°C oven for 2 to 6 hours to digest it until there is no visible big lump.

After cooling, adjust the pH to 6-8 with 10 mol/L NaOH (Do not over acid or over alkali) then constant volume to 4 mL with distilled water. Centrifugation at 16000 rpm for 20 minutes at 25°C (if there is still impurity after centrifugation, it can be removed by filtration).

Take the supernatant for test. (Black substance may be formed in the process, and if it cannot be digested for a long time, it may be carbonized substance. It does not affect the experiment).

2. Cells:

Take 5 million cells, add 1 mL of Extract solution, boil or oven at 110°C for 2 to 6 hours to digest to transparent state.

After cooling, adjust the pH to 6-8 with 10 mol/L NaOH (Do not over acid or over alkali) then constant volume to 2 mL with distilled water. Centrifugation at 16000 rpm for 20 minutes at 25°C (if there is still impurity after centrifugation, it can be removed by filtration).

Take the supernatant for test.

II. Determination procedure

1. Dilute the standard solution to 7.5, 3.75, 1.875, 0.938, 0.469, 0.234, 0.117 µg/mL.

2. Preheat spectrophotometer for 30 minutes, adjust wavelength to 560 nm and set zero with distilled water.

3. Add reagents as the following table.

Reagent (µL)	Blank tube (B)	Test tube (T)	Standard tube (S)
Sample	-	200	-
Standard	-	-	200
Reagent I	200	200	200
Mix well and leave it at room temperature for 20 minutes.			
Reagent II	200	200	200
distilled water	600	400	400

Mix well, incubate at 60°C for 20 minutes, let stand at room temperature for 15 minutes. Use 1mL glass cuvette to measure the OD value of Blank tube, Test tube and Standard tube respectively at 560 nm, and record it as A_B , A_T and A_S .

III. Calculation

1. Making of standard curve.

When making the standard curve, the concentration of the standard solution is taken as the x-axis, and the $\Delta A_S (\Delta A_S = A_S - A_B)$ is taken as the y-axis. The linear equation $y = kx + b$ is obtained. Take $\Delta A_T (\Delta A_T = A_T - A_B)$ to the equation to acquire x.

1. Calculation of hydroxyproline content:

(1) calculated according to the fresh weight of the sample:

$$\text{Tissue hydroxyproline content } (\mu\text{g/g fresh weight}) = x \times V_S \div (W \times V_S \div V_{TE}) = 4x \div W.$$

(2) calculated according to the sample protein concentration:

$$\text{Tissue hydroxyproline content } (\mu\text{g/mg prot}) = x \times V_S \div (C_{pr} \times V_S) = x \div C_{pr}.$$

(3) calculated according to the number of bacteria or cells:

$$\text{Cell hydroxyproline content } (\mu\text{g}/10^4 \text{ cell}) = x \times V_S \div (N \times V_S \div V_{CE}) = 2x \div N.$$

V_S : Volume of added sample, 0.2 mL;

V_{TE} : Volume of tissue extract solution, 4 mL;

V_{CE} : Volume of cell extract solution, 2 mL;

W: Fresh weight of sample, g;

N: Number of cells, 10^4 as a unit;

Cpr: Concentration of sample protein, mg/mL.

Note

1. If the OD value is greater than 1.0, the sample should be diluted properly and then determined. Pay attention to multiply the dilution multiple in the calculation formula.
2. The reagent has certain toxicity. Please take protective measures during operation to prevent inhalation or contact with skin.
3. When calculating according to the sample protein concentration, the protein in the sample itself needs to be extracted separately and determined.

Experimental examples:

0.2 g of mouse leg is taken for sample treatment, and the supernatant is taken out and operated according to the determination steps. $\Delta A = A_T - A_B = 0.243 - 0.022 = 0.221$, and the standard curve $y = 0.0629x + 0.0051$ is brought in, and $x = 3.432$ is calculated.

The content of hydroxyproline in tissue ($\mu\text{g/g}$ mass) = $4x \div W = 4 \times 3.432 \div 0.2 = 68.64 \mu\text{g/g}$ mass.

Recent Product Citations:

[1] Litong Fan, Jiaqing Chen, Yanmeng Tao, et al. Enhancement of the chondrogenic differentiation of mesenchymal stem cells and cartilage repair by ghrelin. Journal of Orthopaedic Research. January 2019; (IF3.043)

Related Products:

NA0744/NA0502	Glutamic-pyruvic Transaminase (GPT) Activity Assay Kit
NA0743/NA0501	Glutamic-oxalacetic Transaminase (GOT) Activity Assay Kit
NA0845/NA0603	Proline (PRO) Content Assay Kit
NA0742/NA0500	Amino Acid (AA) Content Assay Kit
NA0856/NA0614	Cysteine (Cys) Content Assay Kit
NA0741/NA0499	Glutamic Acid (Glu) Content Assay Kit

Technical Specifications:

Detection Limit: $0.051 \mu\text{g/mL}$

Linear Range: $0.117 - 7.5 \mu\text{g/mL}$

